

DECLARATION OF CONFORMITY

Manufacturer: DITEL - Diseños y Tecnología S.A. Address of the manufacturer: Xarol, 6 B (Pol.Ind. Les Guixeres)

08915 Badalona

ESPAÑA

Declares that the product with name: Digital panel meter

Models/Types: JR-P / JR20-P

conforms to following directives:

2014/30/EU Electromagnetic Compatibility

2014/35/EU Low Voltage

2011/65/EC Restriction of the use of certain hazardous

substances in electrical and electronic

equipment

Standards applied:

SAFETY

UNE EN 61010-1:2011 Safety requirements for electrical equipment,

for measurement, control and laboratory use.

Insulation type:

Overvoltage category:

Pollution degree:

Basic insulation
Category II
Degree 2

EMC

EN 61326-1:2013 (Ratified by AENOR) Electrical equipment, for measurement, control

and laboratory use - EMC requirements -

Part 1: General requirements.

Immunity: Industrial electromagnetic environment

Emission: Class B

Signature of Manufacturer

Date: 23-09-2019 Signed: Albert Grau Position: Technical Director